

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

OCT 06 2009

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130130

Name JEREMY CONKLIN ID# 6715 Date 09.31.09
10.01.09

Agency VPD Phone # 835 4560

Instrument Location VPD Jail

Alco S/N 70729 Target Value .081 High Pressure 1400 psi

Alco Test Value Average .081 ✓
1st Alco .081 ✓
2nd Alco

Signature [Signature] 271

(OVER)

BMB 10/13/09

- I, Nita J. Bolz, after being first duly sworn, depose and state as follows:
(Do Not write in the area below)
- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
 - (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
 - (3) I am the Scientific Director of the State Breath Alcohol Program.
 - (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
 - (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
 - (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 22nd day of Oct, 2009.

[Signature]
(Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130130

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130130

OCTOBER 01, 2009

OPERATOR'S NAME:
CONKLING/JEREMY/W
OPERATOR'S NUMBER: 6715
SUBJECT'S LAST NAME:

UOC
SUBJECT'S FIRST NAME/MI :

UOC
O.L. #: UOC
DEPT/AGENCY: UDZ1
CRIME/REPORT: UOC
TEST TYPE: U
ALCO TARGET VALUE: .081
ALCO S/N: 70729

--- BREATH ANALYSIS ---

.081 ADJUSTED FOR 30.09 in
ALCO TARGET 11:59
BLANK TEST .000 11:59
INTERNAL STANDARD VERIFIED 11:59
ALCO TV 30.09 in .081 11:59
BLANK TEST .000 12:00
SUBJECT SAMPLE .000 12:00
BLANK TEST .000 12:01
ALCO TV 30.09 in .081 12:02
BLANK TEST .000 12:02

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130130

OCTOBER 01, 2009
TIME 11:54

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
MEMBERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 40c
BUBBLET: 30.09 in
FLOW DETECTOR: OKAY
POP
MOTOR SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

! " # \$ % ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z \ ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z () * +