MAR 0 8 2011

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130130

Supervisor/Operator Performing the Verification Procedure:

A

Name Jeremy Conkling ID# 6715 Date 03.02.11

Agency VPD Phone # 835-4560

Instrument Location VPD Jail

B

Alco S/N 57023 Target Value .077 High Pressure 200

Alco Test Values .076 (1st Alco) .077 (2nd Alco)

Signature 271

(OVER)

BMB 3/6/11

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.

(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.

(3) I am the Scientific Director of the State Breath Alcohol Program.

(4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.

(5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 31 day of MAR ~~2010~~ 2011

(Notary Seal Stamp)

Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office

STATE OF ALASKA SCIENTIFIC CRIME DETECTION LABORATORY

DANIEL L. NICHOLAS Commission # 116780 NOTARY PUBLIC My Commission Expires With Office STATE OF ALASKA

5

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130130

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130130

MARCH 02, 2011 BMB 3/6/11

OPERATOR'S NAME:
CONKLING/JEREMY/W
OPERATOR'S NUMBER: 6715
SUBJECT'S LAST NAME:
VOC
SUBJECT'S FIRST NAME/MI :
VOC
D.L. #: VOC
DEPT/AGENCY: VOZ1
CASE/REPORT: VOC
TEST TYPE: V
ALCO TARGET VALUE: .077
ALCO S/N: 57023

— BREATH ANALYSIS —

.077 ADJUSTED FOR 29.98 in

ALCO TARGET	.077	11:50
BLANK TEST	.000	11:51
INTERNAL STANDARD	VERIFIED	11:51
ALCO TV 29.98 in	.076	11:51
BLANK TEST	.000	11:52
SUBJECT SAMPLE	.000	11:53
BLANK TEST	.000	11:53
ALCO TV 29.98 in	.077	11:54
BLANK TEST	.000	11:55

BMB 3/6/11

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130130

MARCH 02, 2011
TIME 11:56

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	41c
BAROMETER:	29.98 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
'"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~