


VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUL 27 2011

Supervisor/Operator Performing the Verification Procedure:		Datamaster cdm S/N <u>130131</u>	
Name	<u>Shayne Calt</u>	ID#	<u>4559</u> Date <u>6.13.11</u>
A Agency	<u>AST</u>	Phone #	<u>733-2258</u>
Instrument Location		<u>Talkeetna AST</u>	
B Alco S/N	<u>D119594</u>	Target Value	<u>0.078</u> ^{amb} <u>0.77</u> _{8/11/11} High Pressure <u>300</u>
Alco Test Value Average		<u>.076</u>	<u>.076</u>
		1 st Alco	2 nd Alco
Signature			
(OVER)			

amb
8/11/11

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 18 day of Aug, 2011

NJB


Daniel L. Nicholas Nikki Roth
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N 130131

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130131

JUNE 13, 2011

OPERATOR'S NAME:

CALT SHAYNE M

OPERATOR'S NUMBER: 1000

SUBJECT'S LAST NAME:

NEW

SUBJECT'S FIRST NAME (MI):

NER

D.L. #: U

DEPT/AGENCY: TARA

CASE/REPORT: U

TEST TYPE: U

ALCO TARGET VALUE: 0.070

ALCO S/N: D119531

--- BREATH ANALYSIS ---

0.078 ADJUSTED FOR 29.54 in
ALCO TARGET 0.077 13:36
BLANK TEST 0.000 13:37
INTERNAL STANDARD VERIFIED 13:37
ALCO TU 29.54 in 0.076 13:37
BLANK TEST 0.000 13:38
SUBJECT SAMPLE 0.000 13:38
BLANK TEST 0.000 13:39
ALCO TU 29.54 in 0.076 13:40
BLANK TEST 0.000 13:40

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130131

JUNE 13, 2011

TIME 13:34

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 29.54 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./:;<=>?@ABCDEF
HIJKLMNPOQRSTUVWXYZ[\]^_`abcd efghijklmno
pqrstuvwxyz{|}~