

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

JUL 07 2009

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130131</u> ✓
Name <u>Shayne Carl</u>		ID# <u>4559</u> Date <u>4-5-09</u> ✓
A	Agency <u>AST</u>	Phone # <u>733-2258</u>
Instrument Location <u>Talkeetna AST</u> ✓		
Alco S/N <u>X124665</u> ✓ Target Value <u>.079</u> ✓ High Pressure <u>500</u>		
B	Alco Test Value Average <u>.079</u> ✓	<u>.079</u> ✓
	1 st Alco	2 nd Alco
Signature <u>[Signature]</u>		<u>8/15/09</u>
(OVER)		

BMB
7/14/09


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

(Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 10th day of April, 2009.
[Signature] (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130131

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130131

APRIL 05, 2009 ✓

OPERATOR'S NAME: ✓
CALT SHAYNE M
OPERATOR'S NUMBER: 4559
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
U

O.L. #: U
DEPT/AGENCY: TALA
CASE/REPORT: U
TEST TYPE: U
ALCO TARGET VALUE: .079 ✓
ALCO S/N: X124665

BREATH ANALYSIS

.079 ADJUSTED FOR 29.80 in ✓
ALCO TARGET .078 16:21
BLANK TEST .000 16:22
INTERNAL STANDARD VERIFIED 16:22
ALCO TU 29.80 in .079 16:22
BLANK TEST .000 ✓ 16:23
SUBJECT S/N .000 16:24
BLANK TEST .000 16:24
ALCO TU 29.80 in .079 ✓ 16:25
BLANK TEST .000 16:25

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130131

APRIL 05, 2009 ✓
TIME 16:19

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 43c
BAROMETER: 29.80 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST
!#\$%&'()*+,-./0123456789:;<=>?@ABCDEF
GHIJKLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmnop
qrstuvwxy{|}~■