

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130131 ✓
Name Shayne Calt ID# 4559 Date 11-13-08 ✓
A Agency AST Phone # 733-2256
Instrument Location Talkeetna AST

B Alco S/N X124665 ✓ Target Value .079 ✓ High Pressure 500 ✓
Alco Test Value Average .078 ✓ .078 ✓
1st Alco 2nd Alco
Signature [Signature] ✓ CSB
12/5/08

Under the Alaska Rules of Evidence, I certify that, (OVER)

(a) I am ^(Do Not write in this area please) Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Shayne M Calt, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,

[Signature]

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

12/5/08

Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130131 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130131

NOVEMBER 13, 2008 ✓ *COB 12/15/08*

OPERATOR'S NAME: CALT SHAYNE M.
OPERATOR'S NUMBER: 4559
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI:
U
ILL. #: U
DEPT/AGENCY: TALA
CASE/REPORT: U
TEST TYPE: U
ALCO TARGET VALUE: .079 ✓
ALCO S/N: X124665

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 29.70 in		
ALCO TARGET	.078	15:47
BLANK TEST	.000	15:48
INTERNAL STANDARD	VERIFIED	15:48
ALCO TV 29.70 in	.078 ✓	15:48
BLANK TEST	.000	15:49
SUBJECT SAMPLE	.000	15:50
BLANK TEST	.000	15:51
ALCO TV 29.70 in	.078 ✓	15:51
BLANK TEST	.000	15:52

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130131

NOVEMBER 13, 2008 ✓
TIME 15:46

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01

HEATERS
SAMPLE CHAMBER: 50c
BREATH TUBE: 41c

BAROMETER: 29.70 in

FLOW DETECTOR: OKAY

PUMP
HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST
!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKL MNOPQRSTUVWXYZ\]^_`~:;abcde fghijklmno
pqrstuvwxyz{|}~@