

# VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUN 22 2010

Supervisor/Operator Performing the Verification Procedure:		Datamaster cdm S/N <u>130131</u> ✓	
Name <u>Shayne Calt</u>		ID# <u>4559</u> ✓	Date <u>5-14-10</u> ✓
A	Agency <u>AST</u>	Phone # <u>733-2256</u>	
Instrument Location <u>Talkeetna</u>			
Alco S/N <u>31343</u> ✓		Target Value <u>.084</u> ✓	High Pressure <u>400</u>
B	Alco Test Value Average <u>.083</u> ✓		<u>.082</u> ✓
		1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco
Signature <u>[Signature]</u>			
(OVER)			

*[Handwritten initials]*  
6/29/10


I, Nita J. Bolz, after being first duly sworn, depose and state as follows: (Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 24<sup>th</sup> day of June, 2010.

[Signature]  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N 130131

(CONTINUED FROM FRONT PAGE)

**C** TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130131

MAY 14, 2010

OPERATOR'S NAME: *BMB 6/23/10*  
CALT SHAYNE M  
OPERATOR'S NUMBER: 4559  
SUBJECT'S LAST NAME:  
VERIFICATION/FAL  
SUBJECT'S FIRST NAME/MI :  
U  
O.L. #: U  
DEPT/AGENCY: TALO  
CASE/REPORT: U  
TEST TYPE: U  
ALCO TARGET VALUE: .084  
ALCO S/N: 31343

--- BREATH ANALYSIS ---

.084 ADJUSTED FOR 29.65 in		
ALCO TARGET	.083	12:41
BLANK TEST	.000	12:41
INTERNAL STANDARD	VERIFIED	12:42
ALCO TU 29.65 in	.083	12:42
BLANK TEST	.000	12:43
SUBJECT SAMPLE	.000	12:43
BLANK TEST	.000	12:44
ALCO TU 29.65 in	.082	12:44
BLANK TEST	.000	12:45

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130131

MAY 14, 2010  
TIME 12:38

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01

HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 41c

BAROMETER: 29.65 in

FLOW DETECTOR: OKAY

PUMP  
HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST  
!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEF6  
HIJKLHNDOPQRSTUWXYZ\]^\_`abcde{ghijklmno  
pqrstuvwxyz{|}~