

VERIFICATION OF CALIBRATION REPORT JUN 13 2008

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: <u>DataMaster cdm S/N 130132</u>	
Name <u>Doreen Valadez</u>	ID# <u>3423</u> Date <u>6/11/08</u>
Agency <u>Seward Police Dept.</u>	Phone # <u>224-3338</u>
Instrument Location <u>Seward Community Jail</u>	
Alco S/N <u>X124623</u>	Target Value <u>0.078</u> High Pressure <u>1100</u>
Alco Test Value Average <u>0.079</u>	<u>0.080</u>
1 st Alco	2 nd Alco
Signature <u>Doreen R. Valadez</u>	<u>COB</u> <u>6/27/08</u>

(OVER)


Under the Alaska Rules of Evidence, I certify that, in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Doreen R Valadez, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



Jeanne Swartz

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

6/30/08
Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 13 0132

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130132

JUNE 11, 2008 *COS 6/24/08*

OPERATOR'S NAME:
VALADEZ D R

OPERATOR'S NUMBER: 3423

SUBJECT'S LAST NAME:
VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :
X

D.L. #: X

DEPT/AGENCY: SMD1

CASE/REPORT: X

TEST TYPE: V

ALCO TARGET VALUE: .078

ALCO S/N: X124623

--- BREATH ANALYSIS ---

.078 ADJUSTED FOR 30.30 in

ALCO TARGET	.079	17:00
BLANK TEST	.000	17:02
INTERNAL STANDARD	VERIFIED	17:02
ALCO TO 30.30 in	.091	17:02
BLANK TEST	.000	17:03
SUBJECT SAMPLE	.000	17:03
BLANK TEST	.000	17:04
ALCO TO 30.30 in	.050	17:05
BLANK TEST	.000	17:05

Del

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130132

JUNE 11, 2008
TIME 17:05

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 03/29/00

HEATERS

SAMPLE CHAMBER: 43c

BREATH TUBE: 44c

BAROMETER: 30.30 in *Del*

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJKL
MNPQRSTUVWXYZ[\]^_`abcdefghijklmnopqrstuvwxyz0123456789:;<=>?@ABCDEFGHIJKL