

MAY 04 2009

VERIFICATION OF CALIBRATION REPORT
of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130133 ✓

Name BRYAN BARLOW ID# 4548 Date 4/26/09 ✓

A Agency ALASKA STATE Troopers Phone # 907-783-0972

Instrument Location AST - Girdwood ✓

B Alco S/N x124671 ✓ Target Value .077 ✓ High Pressure 400

Alco Test Value Average .081 ✓ .082 ✓
1st Alco 2nd Alco

Signature Bryan Barlow ✓

CUB
5/13/09

(OVER)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:
(Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 14 day of May 2009.

Carolyn M. Notland (Notary Seal Stamp)
Carolyn M. Notland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130133 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130133

APRIL 26, 2009 ✓

OPERATOR'S NAME: *CBS Sliver*
BARLOW BRYAN W
OPERATOR'S NUMBER: 4548
SUBJECT'S LAST NAME:
OF CALIBRATION
SUBJECT'S FIRST NAME/NI :
VERIFICATION
O.L. #: 1234567
DEPT/AGENCY: GDMO
CASE/REPORT: 09-123456
TEST TYPE: U
ALCO TARGET VALUE: .077 ✓
ALCO S/N: X124671

BREATH ANALYSIS

.077 ADJUSTED FOR 30.98 in
ALCO TARGET .079 21:21
BLANK TEST .000 21:22
INTERNAL STANDARD VERIFIED 21:22
ALCO TU 31.00 in .081 21:22
BLANK TEST .000 21:23
SUBJECT SAMPLE *CBS Sliver* .000 21:24
BLANK TEST .000 21:24
ALCO TU 31.00 in .082 21:25
BLANK TEST .000 21:26

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130133

APRIL 26, 2009 ✓
TIME 21:15

DIAGNOSTIC CHECK

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 48c
BREATH TUBE: 43c
BAROMETER: 31.00 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@BCDEFG
HIJKL MNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~■