OCT 2 6 2010

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130134

Supervisor/Operator Performing the Verification Procedure:

A

Name DEREK A. MCCARRIGAN ID# 5420 Date 10/1/10

Agency KETCHIKAN POLICE DEPARTMENT Phone # 907-225-6631

Instrument Location KTN. POLICE DEPT., 361 MAIN ST., KTN. AK 99901

B

Alco S/N X301765 Target Value .079 High Pressure 400

Alco Test Values .081 (1st Alco) .081 (2nd Alco)

Signature

(OVER)

BMB 12/8/10

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.

(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.

(3) I am the Scientific Director of the State Breath Alcohol Program.

(4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.

(5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 16th day of DEC, 2010.

(Notary Seal Stamp)

Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office

STATE OF ALASKA SCIENTIFIC CRIME DETECTION LABORATORY

DANIEL L. NICHOLAS Commission # 116780 NOTARY PUBLIC My Commission Expires With Office STATE OF ALASKA

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130134 BMB 12/8/10

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130134

OCTOBER 01, 2010 BMB 12/8/10

OPERATOR'S NAME:
MCGARRIGAN DEREK A
OPERATOR'S NUMBER: 5420
SUBJECT'S LAST NAME:
VOC
SUBJECT'S FIRST NAME/MI :
VOC
O.L. #: 1234567
DEPT/AGENCY: KET1
CASE/REPORT: 10-1
TEST TYPE: V
ALCO TARGET VALUE: .079
ALCO S/N: X301765

— BREATH ANALYSIS —

.079 ADJUSTED FOR 30.07 in

ALCO TARGET	.079	15:36
BLANK TEST	.000	15:37
INTERNAL STANDARD	VERIFIED	15:37
ALCO TV 30.07 in	.081	15:38
BLANK TEST	.000	15:38
SUBJECT SAMPLE	.000	15:39
BLANK TEST	.000	15:39
ALCO TV 30.07 in	.081	15:40
BLANK TEST	.000	15:40

BMB 12/8/10

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130134

OCTOBER 01, 2010
TIME 15:35

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/08/09
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	41c
BAROMETER:	30.07 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~