

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

MAY 28 2010

67-10
CW

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 78855 130135

Name Trp. J. Santora ID# 2276 Date 5/25/10

A Agency Alaska State Troopers Phone # 656-1233

Instrument Location Galena Police Department

B Alco S/N 78855 Target Value .081 High Pressure 1225

Alco Test Values .080 .080
1st Alco 2nd Alco

Signature J. Santora

(OVER)

(Do Not write in the area below)

- I, Nita J. Bolz, after being first duly sworn, depose and state as follows:
- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
 - (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
 - (3) I am the Scientific Director of the State Breath Alcohol Program.
 - (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
 - (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
 - (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 2nd day of July, 2010.

Carolyn M. Noland (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N _____ 130135 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST



ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130135 ✓

MAY 25, 2010 ✓


OPERATOR'S NAME: *COB 6/24/10*
SANTORA, JEANNINE/L
OPERATOR'S NUMBER: 2276
SUBJECT'S LAST NAME:
VERIFICATION
SUBJECT'S FIRST NAME/MI :
OF CAL
O.L. #: 0
DEPT/AGENCY: GALO
CASE/REPORT:
TEST TYPE: U
ALCO TARGET VALUE: .081 ✓
ALCO S/N: 78855 ✓

--- BREATH ANALYSIS ---

.081 ADJUSTED FOR 30.87 in ✓
ALCO TARGET .081 19:44
BLANK TEST .000 19:46
INTERNAL STANDARD VERIFIED 19:46
ALCO TV 30.87 in .080 ✓ 19:46
BLANK TEST .000 19:47
SUBJECT SAMPLE *COB 6/24/10* .000 19:48
BLANK TEST .000 19:48
ALCO TV 30.87 in .080 ✓ 19:49
BLANK TEST .000 19:49



DIAGNOSTIC CHECK



ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130135 ✓

MAY 25, 2010
TIME 19:35

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 48c
BREATH TUBE: 43c
BAROMETER: 30.87 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@BCDEFG
HIJKLMNPOQRSTUVWXYZ[\]^_`abcedefghijklmnop
qrstuvwxyz{|}~■