

AUG 31 2011

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130141</u>	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>Scott Gardner</u>	ID# <u>6222</u> Date <u>29 Aug 2011</u>
A Agency <u>673 Security Forces Squadron JBR4</u>	Phone # <u>907-552-3537</u>
Instrument Location <u>Bldg 8517 20th St Joint Base Elmendorf-Richardson, AK</u>	
B Alco S/N <u>4A</u> Target Value <u>.080</u> High Pressure <u>500 psi</u>	
Alco Test Values	<u>.080</u> <u>.079</u>
	1st Alco 2nd Alco
Signature <u><i>Scott Gardner</i></u>	

(OVER)

BMB
9/2/11

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 2 day of Sept, 2010. *(Signature)*

Nikki Roth (Notary Seal Stamp)
~~Daniel L. Nicholas~~ *Nikki Roth*
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130141

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130141
AUGUST 29, 2011
OPERATOR'S NAME:
GARDNER/SCOTT/E
OPERATOR'S NUMBER: 6222
SUBJECT'S LAST NAME:
N/A
SUBJECT'S FIRST NAME/MI :
N/A
O.L. #: VERIFICATION OF CALIBRATION
DEPT/AGENCY: JBR4
CASE/REPORT: N/A
TEST TYPE: V
ALCO TARGET VALUE: .000
ALCO S/N: 4A

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 30.01 in
ALCO TARGET .000 00:21
BLANK TEST .000 00:22
INTERVAL STANDARD VERIFIED 00:22
ALCO TV 30.01 in .000 00:23
BLANK TEST .000 00:23
SUBJECT SAMPLE .000 00:24
BLANK TEST .000 00:25
ALCO TV 30.01 in .079 00:25
BLANK TEST .000 00:26

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130141

AUGUST 29, 2011
TIME 08:18

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 40c
BREATH TUBE: 43c
BAROMETER: 30.01 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKLMN0PQRSTUWXYZ\]^_`~"abcde fghijklmno
pqrstu vwx yz{|}~■

