FEB 0 7 ~~2010~~ 2011 3/14/11

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130142

Supervisor/Operator Performing the Verification Procedure:

A

Name Lynne Gralewski-Baugus ID# 6787 Date 31 Jan 11

Agency Joint Base Elemendorf-Richardson Phone # 907-552-4000

Instrument Location Bldg 656 A St JBER, AK 99506

B

Alco S/N X172946 Target Value .086 High Pressure 500 psi

Alco Test Values .087 (1st Alco) .087 (2nd Alco)

Signature LM Gralewski-Baugus

(OVER)

BMB 3/16/11

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
(3) I am the Scientific Director of the State Breath Alcohol Program.
(4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
(5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 14th day of MAR, ~~2010~~ 2011

(Notary Seal Stamp)

Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office

STATE OF ALASKA SCIENTIFIC CRIME DETECTION LABORATORY

DANIEL L. NICHOLAS Commission # 116780 NOTARY PUBLIC My Commission Expires With Office STATE OF ALASKA

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130142

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130142

JANUARY 31, 2011 BMB 3/6/11

OPERATOR'S NAME:
GROLEWSKI-BAUGUS/L/WHE/M
OPERATOR'S NUMBER: 6737
SUBJECT'S LAST NAME:
VERIFICATION OF CALIBRATION
SUBJECT'S FIRST NAME/MI :
UOC
O.L. #: U
DEPT/AGENCY: EDF4
CASE/REPORT: U
TEST TYPE: U
ALCO TARGET VALUE: 0.086
ALCO S/N: X172945

--- BREATH ANALYSIS ---

.086 ADJUSTED FOR 29.67 in

ALCO TARGET	.085	07:31
BLANK TEST	.000	07:31
INTERNAL STANDARD	VERIFIED	07:32
ALCO TV 29.67 in	.087	07:32
BLANK TEST	.000	07:33
SUBJECT SAMPLE	.000	07:33
BLANK TEST	.000	07:34
ALCO TV 29.68 in	.087	07:35
BLANK TEST	.000	07:36

BMB 3/6/11

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130142

JANUARY 31, 2011
TIME 07:26

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	03/29/00
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	44c
BAROMETER:	29.67 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~