

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

NOV 02 2009

Supervisor/Operator Performing the Verification Procedure:	
DataMaster cdm S/N <u>130142</u> ✓	
Name <u>Scott E. Gardner</u> ID# <u>6222</u> Date <u>29 Oct 09</u> ✓	
A	Agency <u>Directorate of Emergency Services</u> Phone # <u>384-3888</u>
	Instrument Location <u>Bldg #656, A St, Ft Richardson, AK 99505</u>
B	Alco S/N <u>X124656</u> ✓ Target Value <u>.080</u> ✓ High Pressure <u>1175 psi</u>
	Alco Test Values <u>.077</u> ✓ <u>.080</u> ✓ 1 st Alco 2 nd Alco
	Signature <u>Scott E. Gardner</u> ✓
	(OVER)

SMB
11/14/09

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 30th day of Dec, 2009.

Carolyn M. Noland (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130142 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
 DATAMASTER cdm 130142

OCTOBER 29, 2009

OPERATOR'S NAME:
 BARNER/SCOTT/E

OPERATOR'S NUMBER: 8222

SUBJECT'S LAST NAME:
 SUBJECT'S FIRST NAME: N/A

DL# N/A

DEPT/AGENCY: FR-4

CASE/ARREST#: N/A

TEST TYPE: N/A

ALCOHOL TEST VALUE: 0.00 ✓

--- BREATH ANALYSTS ---

ALCOHOL TEST	0.00	14147
INTERNAL STANDARD	VERIFIED	14148
ALCOHOL TEST	0.00	14149
INTERNAL STANDARD	VERIFIED	14150
ALCOHOL TEST	0.00	14151
INTERNAL STANDARD	VERIFIED	14152

Handwritten notes: BMB 11/4/09, CCB x 1241656 11/12/09

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
 DATAMASTER cdm 130142 ✓

OCTOBER 29, 2009 ✓

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 03/29/08

STATUS:

WATER CHAMBER: 45C

BREATH TUBE: 45C

BAROMETER: 29.75 in

FLOW DETECTOR: OKAY

PUMP HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST: OKAY

ALCOHOL TEST VALUE: 0.00 ✓

INTERNAL STANDARD: VERIFIED ✓