VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130143

Supervisor/Operator Performing the Verification Procedure:

A

Name LT DAVE MCKILLICAN ID# 5210 Date 5/9/2008

Agency FT WAINWRIGHT POLICE Phone # 353-7531

Instrument Location 3028 MONTGOMERY ROAD FT WAINWRIGHT AK

B

Alco S/N X 172958 Target Value .082 High Pressure 1000 PSI

Alco Test Values .085 (1st Alco) .084 (2nd Alco)

Signature [signature] 06/13/08

(OVER)

(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that,

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, William McKillican, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

[signature] 6/16/08

Jeanne Swartz Date
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

STATE OF ALASKA SCIENTIFIC CRIME DETECTION LABORATORY

BT9 06/03

MAY 15 2009

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130143

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130143

MAY 09, 2008 CDB 6/13/08

OPERATOR'S NAME:
MCKILLICAN/WILLIAM/O
OPERATOR'S NUMBER: 5210
SUBJECT'S LAST NAME:
VOC
SUBJECT'S FIRST NAME/MI :
VOC
O.L. #: 123456
DEPT/AGENCY: FA14
CASE/REPORT: 08-123456
TEST TYPE: T
ALCO TARGET VALUE: .082
ALCO S/N: X172958

--- BREATH ANALYSIS ---

.082 ADJUSTED FOR 29.44 in

ALCO TARGET	.080	14:04
BLANK TEST	.000	14:05
INTERNAL STANDARD	VERIFIED	14:05
ALCO TU 29.44 in	.085	14:05
BLANK TEST	.000	14:06
SUBJECT SAMPLE	.080 CDB 6/13/08	14:07
BLANK TEST	.000	14:07
ALCO TU 29.44 in	.084	14:08
BLANK TEST	.000	14:09

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130143

MAY 09, 2008
TIME 14:01

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	29.44 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~