



VERIFICATION OF CALIBRATION REPORT
of DataMaster cdm Breath Test Instrument **AUG 27 2009**
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:			DataMaster cdm S/N 130144 ✓
Name: Christopher Chapman	ID# 6390	Date 24 Aug 2009 ✓	
A Agency 3 rd Security Forces Squadron	Phone # (907)552-0176		
Instrument Location: Law Enforcement desk, Elmendorf AFB, AK ✓			
B			
Alco S/N X172959 ✓	Target Value .081 ✓	High Pressure 600psi	
Alco Test Values	0.081 ✓ 1 st Alco	0.080 ✓ 2 nd Alco	
Signature			COB 8/31/09
(OVER)			


(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 4th day of Sept., 2009.

 (Notary Seal Stamp)
 Carolyn M. Noland
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130144

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130144

AUGUST 24, 2009

OPERATOR'S NAME: CHAPMAN CHRISTOPHER M
OPERATOR'S NUMBER: 6390
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
VOC

O.L. #: 1
DEPT/AGENCY: EDF4
CASE/REPORT:
TEST TYPE: U
ALCO TARGET VALUE: .081
ALCO S/N: N172959

— BREATH ANALYSIS —

.081 ADJUSTED FOR 29.70 in
ALCO TARGET .080 13:52
BLANK TEST .080 13:53
INTERNAL STANDARD VERIFIED 13:53
ALCO TU 29.70 in .081 13:54
BLANK TEST .080 13:54
SUBJECT SAMPLE .080 13:55
BLANK TEST .080 13:56
ALCO TU 29.70 in .080 13:56
BLANK TEST .080 13:57

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130144

AUGUST 24, 2009
TIME 13:58

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 01/08/09
HEATERS
SAMPLE CHAMBER: 50c
BREATH TUBE: 41c
BAROMETER: 29.70 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEF
HIJKLMNPOQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxy z{|}~