

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130144 owo
12/19/10

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130144

NOVEMBER 10, 2010

OPERATOR'S NAME: GARDNER//SCOTT/E
OPERATOR'S NUMBER: 6222
SUBJECT'S LAST NAME: N/A
SUBJECT'S FIRST NAME/MI: N/A
O.L. #: VERIFICATION OF CAL
DEPT/AGENCY: EDF4
CASE/REPORT: N/A
TEST TYPE: U
ALCO TARGET VALUE: .080
ALCO S/N: 2T

--- BREATH ANALYSIS ---

.080 ADJUSTED FOR 29.85 in
ALCO TARGET .079 14:42
BLANK TEST .000 14:43
INTERNAL STANDARD VERIFIED 14:43
ALCO TV 29.86 in .081 14:43
BLANK TEST .000 14:44
SUBJECT SAMPLE .000 14:44
BLANK TEST .000 14:45
ALCO TV 29.85 in .079 14:45
BLANK TEST .000 14:46

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130144

NOVEMBER 10, 2010
TIME 14:48

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 01/08/09
HEATERS
SAMPLE CHAMBER: 50c
BREATH TUBE: 41c
BAROMETER: 29.86 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF
GHIJKLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmnop
qrstuvwxy{|}~!@