

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

MAR 10 2010

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130172 ✓

Name TIM TUCKWOOD ID# 3446 Date 3-10-10 ✓

A Agency Alaska State Troopers Phone # 895-4800

Instrument Location TOK AST

B Alco S/N 130172 ✓ Target Value 083 High Pressure 700 psi

3151020
X173017

Alco Test Value Average 081 ✓ 081 ✓
1st Alco 2nd Alco

Signature Tim Tuckwood ✓

(OVER)

BMB
3/24/10

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 5th day of April, 2010.

Carolyn M. Noland (Notary Seal Stamp)

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

131172 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130172

MARCH 10, 2010 ✓

OPERATOR'S NAME:

TUCKWOOD TIM

OPERATOR'S NUMBER: 3446

SUBJECT'S LAST NAME:

UBIF

SUBJECT'S FIRST NAME/MI :

OF CAL

O.L. #: 12345

DEPT/AGENCY: TOKO

CHIEF/REPORT: 09076

TEST TYPE: U

ALCO TARGET VOL%: 0.00

ALCO S/N: XT73017 ✓

— BREATH ANALYSIS —

ALCO ADJUSTED FOR 27.85 in

ALCO TARGET	.077	13:00
BLANK TEST	.000	13:01
INTERNAL STANDARD	VERIFIED	13:01
ALCO TV 27.85 in	.061 ✓	13:02
BLANK TEST	.000	13:02
SUBJECT SAMPLE	.000	13:03
BLANK TEST	.000	13:04
ALCO TV 27.85 in	.061 ✓	13:04
BLANK TEST	.000	13:05

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130172 ✓

MARCH 10, 2010 ✓
TIME 12:57

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PRINTER: OKAY
SOFTWARE DATE: 01/08/09
METERS
SAMPLE CHAMBER: 49c
TUBE: 41c
BAROMETER: 27.87 in
FLOW DETECTOR: OKAY
MOTOR SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~