

MAR 11 2011

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster cdm Breath Test Instrument*

*State of Alaska*

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

DataMaster cdm S/N 130172 ✓

Supervisor/Operator Performing the Verification Procedure:  
 Name Sgt. Freddie Wells ID# 4246 ✓ Date 3-8-11 ✓  
 A Agency Alaska State Troopers Phone # 907-883-5111  
 Instrument Location TOK, Alaska

---

B Alco S/N X173017 ✓ Target Value .083 ✓ High Pressure 650  
 Alco Test Values .083 ✓ .085 ✓  
                                     1st Alco                                      2nd Alco  
 Signature [Signature]

(OVER)

AMB  
4/1/11

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]  
 Nita J. Bolz  
 Scientific Director  
 State Breath Alcohol Program

Subscribed and sworn before me this 7<sup>th</sup> day of APRIL, 2010. dm

[Signature] (Notary Seal Stamp)  
 Daniel L. Nicholas  
 Notary Public, State of Alaska  
 Commission Expires with Office

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument  
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130172

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130172

MARCH 08, 2011

OPERATOR'S NAME:  
WELLS/FREDDIE/L  
OPERATOR'S NUMBER: 4246  
SUBJECT'S LAST NAME:

UOC  
SUBJECT'S FIRST NAME/MI :  
UOC

O.L. #: 1234567  
DEPT/AGENCY: TONE  
CASE/REPORT: UOC  
TEST TYPE: V  
ALCO TARGET VALUE: .083  
ALCO S/N: X173017

### — BREATH ANALYSIS —

.083 ADJUSTED FOR 28.55 in  
ALCO TARGET .079 10:23  
BLANK TEST .000 10:24  
INTERNAL STANDARD VERIFIED 10:24  
ALCO TV 28.55 in .083 10:25  
BLANK TEST .000 10:25  
SUBJECT SAMPLE .000 10:26  
BLANK TEST .000 10:27  
ALCO TV 28.55 in .085 10:27  
BLANK TEST .000 10:28

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130172

MARCH 08, 2011  
TIME 10:28

### — DIAGNOSTIC CHECK —

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 01/08/09  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 41c  
BAROMETER: 28.55 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

!@#%&'()\*+,-./0123456789:;<=>?@ABCDEF  
HIJKLMNPOQRSTUVWXYZ[\]^\_`abcde fghi jk l m n o  
p q r s t u v w x y z { | } ~ ¡ ¢