

FEB 06 2008 **VERIFICATION OF CALIBRATION REPORT**

*of Datamaster cdm*

*State of Alaska*

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Datamaster cdm S/N ~~#~~ 30173

Supervisor/Operator Performing the Verification Procedure:

Name Josh Heinbaugh ID# 5828 Date 2-4-08

**A** Agency AST Phone # 822-3263

Instrument Location Glennallen

---

**B** Alco S/N 77198 Target Value 0.083 High Pressure 700

Alco Test Value Average 0.078 0.078

1<sup>st</sup> Alco 2<sup>nd</sup> Alco

Signature [Signature] CSB  
2/21/08

(OVER)

Under the Alaska Rules of Evidence, I certify that,

(Do Not write in the area below)

(a) I am Jeanne Swartz, Criminalist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Josh Heinbaugh, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed



[Signature] 02/22/08

Date

Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

# VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N 130173

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130173

FEBRUARY 04, 2000

OPERATOR'S NAME: HEINDLICH  
OPERATOR'S NUMBER: 5828  
SUBJECT'S LAST NAME: VOC

SUBJECT'S FIRST NAME/INITIAL: VOC

D.L. #: VOC  
DEPT/AGENCY: GNK8  
CASE/REPORT: VOC  
TEST TYPE: U  
ALCO TARGET VALUE: 0.083  
ALCO S/N: 77198

--- BREATH ANALYSIS ---

.083 ADJUSTED FOR 28.30 in  
ALCO TARGET .078 18:24  
BLANK TEST .000 18:25  
INTERNAL STANDARD (VERIFIED) 18:25  
ALCO TU 28.30 in .078 18:26  
BLANK TEST .000 18:26  
SUBJECT SAMPLE .000 18:27  
BLANK TEST .000 18:28  
ALCO TU 28.30 in .078 18:28  
BLANK TEST .000 18:29

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130173

FEBRUARY 04, 2000  
TIME 18:21

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 42c  
BAROMETER: 28.30 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()\*+,-./0123456789:;<=>?@ABCDEF  
GHIJKLMNOPQRSTUVWXYZ[\]^\_`abcde fghi jk lmn  
opqrs tuv wxyz{|}~