

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska


Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130173</u>
A	Name	<u>Heinbaugh</u>	ID# <u>5823</u> Date <u>6-10-08</u> ✓
	Agency	<u>AST</u>	Phone # <u>822-3263</u>
	Instrument Location	<u>Glennallen</u> ✓	
B	Alco S/N	<u>77198</u> ✓	Target Value <u>0.083</u> ✓ High Pressure <u>500</u>
	Alco Test Values	<u>0.078</u> ✓ 1 st Alco	<u>0.079</u> ✓ 2 nd Alco
	Signature	<u>[Signature]</u> COB 6/20/08	

(OVER)

(Do Not write in the area below)
Under the Alaska Rules of Evidence, I certify that;

- (a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.
- (b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Josh Heinbaugh, whose name appears in my records, is qualified to verify the calibration of the above instrument.
- (c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,

[Signature]

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

6/20/08

Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130173

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATA MASTER cdm 130173

JUNE 10, 2008

OPERATOR'S NAME:

HEINBAUGH

OPERATOR'S NUMBER: 5828

SUBJECT'S LAST NAME:

UOC

SUBJECT'S FIRST NAME/MI :

UOC

O.L. #: UOC

DEPT/AGENCY: GNKA

CASE/REPORT: UOC

TEST TYPE: U

ALCO TARGET VALUE: 0.083

ALCO S/N: 77198

--- BREATH ANALYSIS ---

.083 ADJUSTED FOR 28.56 in		
ALCO TARGET	.079	14:42
BLANK TEST	.000	14:43
INTERNAL STANDARD	VERIFIED	14:43
ALCO TV 28.55 in	.078	14:44
BLANK TEST	.000	14:44
SUBJECT SAMPLE	.000	14:45
BLANK TEST	.000	14:45
ALCO TV 28.56 in	.079	14:46
BLANK TEST	.000	14:47

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATA MASTER cdm 130173

JUNE 10, 2008

TIME 14:52

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	48c
BREATH TUBE:	41c
BAROMETER:	28.56 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=?@ABCDEFGHI
HIJKLMNPOQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~