

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster cdm Breath Test Instrument  
State of Alaska*

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

DataMaster cdm S/N 130173 ✓

Supervisor/Operator Performing the Verification Procedure:  
 Name Jessie Hemburg ID# 5828 Date 8-22-10

**A** Agency AST Phone # 907-3263

Instrument Location Glennallen

---

**B** Alco S/N X 127664 Target Value 0.083 High Pressure 500

Alco Test Values .081 <sup>ans</sup> 8/22/10 .082 <sup>ans</sup> 8/22/10  
 1st Alco 2nd Alco

Signature [Signature]

(OVER)

*BMB  
8/22/10*

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]  
 Nita J. Bolz  
 Scientific Director  
 State Breath Alcohol Program

Subscribed and sworn before me this 5<sup>th</sup> day of Oct, 2010.

[Signature] (Notary Seal Stamp)  
 Daniel L. Nicholas  
 Notary Public, State of Alaska  
 Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument  
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

13 0173

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130173

AUGUST 22, 2010

OPERATOR'S NAME:  
HEINBAUGH  
OPERATOR'S NUMBER: 5828  
SUBJECT'S LAST NAME:  
VOC  
SUBJECT'S FIRST NAME/MI :  
VOC

O.L. #: VOC  
DEPT/AGENCY: GNK0  
CASE/REPORT: VOC  
TEST TYPE: V  
ALCO TARGET VALUE: 0.083  
ALCO S/N: X124669

### --- BREATH ANALYSIS ---

.083 ADJUSTED FOR 28.37 in  
ALCO TARGET .078 12:22  
BLANK TEST .000 12:23  
INTERNAL STANDARD VERIFIED 12:23  
ALCO TV 28.37 in .001 12:24  
BLANK TEST .000 12:24  
SUBJECT SAMPLE .000 12:25  
BLANK TEST .000 12:26  
ALCO TV 28.37 in .082 12:26  
BLANK TEST .000 12:27

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130173

AUGUST 22, 2010  
TIME 12:19

### --- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 42c  
BAROMETER: 28.37 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

!@#%&'()\*+,-./0123456789:;<=>?@ABCDEFGHI  
HIJKL MNOPQRSTUVWXYZ[\]^\_`abcdefgijklmno  
pqrstuvwxyz{|}~