

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUN 20 2011

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130174

Name Lowe, Gary ID# 7219 Date 6/16/2011

A Agency Haines Police Phone # 766-2121

Instrument Location Haines P.D.

B Alco S/N X172978 Target Value .075 High Pressure 1500

Alco Test Values .076 .075
1st Alco 2nd Alco

Signature [Handwritten Signature]

(OVER)

[Handwritten initials]
6/20/11

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Handwritten Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 12 day of Aug, 2011.

[Handwritten Signature] (Notary Seal Stamp)
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130174

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130174

JUNE 16, 2011

OPERATOR'S NAME:

LOWE, GARY

OPERATOR'S NUMBER: 7219

SUBJECT'S LAST NAME:

TEST

SUBJECT'S FIRST NAME/MI :

VERIFY

O.L. #: 06152011

DEPT/AGENCY: HNS1

CASE/REPORT: 11-0616

TEST TYPE: U

ALCO TARGET VALUE: .075

ALCO S/N: X172978

--- BREATH ANALYSIS ---

.075 ADJUSTED FOR 30.29 in

ALCO TARGET .075 14:48

BLANK TEST .000 14:49

INTERNAL STANDARD VERIFIED 14:49

ALCO TU 30.27 in .075 14:50

BLANK TEST .000 14:50

SUBJECT SAMPLE .000 14:51

BLANK TEST .000 14:51

ALCO TU 30.27 in .075 14:52

BLANK TEST .000 14:52

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130174

JUNE 16, 2011

TIME 14:16

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 01/08/09

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 41c

BAROMETER: 30.29 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@BCDEFG
HIJKL MNOPQRSTUVWXYZ \]^_` abcdefghijklmno
pqrstuvwxyz{|}~■