

VERIFICATION OF CALIBRATION REPORT SEP 14 2011
of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N <u>130174</u>					
Name <u>LOWE, GARY</u>	ID# <u>7219</u> Date <u>9/9/2011</u>				
A Agency <u>Haines Police</u> Phone # <u>766-2121</u>					
Instrument Location <u>Haines P.D.</u>					
B Alco S/N <u>X172978</u> Target Value <u>.075</u> High Pressure <u>1500</u>					
Alco Test Values	<table style="width: 100%; border: none;"> <tr> <td style="text-align: center; border-bottom: 1px solid black;"><u>.076</u></td> <td style="text-align: center; border-bottom: 1px solid black;"><u>.076</u></td> </tr> <tr> <td style="text-align: center;">1st Alco</td> <td style="text-align: center;">2nd Alco</td> </tr> </table>	<u>.076</u>	<u>.076</u>	1 st Alco	2 nd Alco
<u>.076</u>	<u>.076</u>				
1 st Alco	2 nd Alco				
Signature <u>Gary Lowe</u>	<u>CB</u> <u>9/15/11</u>				

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 7 day of Nov, 2011 AK

Nikki Roth (Notary Seal Stamp)
 Daniel L. Nicholas
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130174

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130174

SEPTEMBER 09, 2011 *COB 9/15/11*

OPERATOR'S NAME: LOWE, GARY
OPERATOR'S NUMBER: 7219
SUBJECT'S LAST NAME: TEST
SUBJECT'S FIRST NAME/MI: VER
O.L. #: 12345
DEPT/AGENCY: HNSI
CASE/REPORT: 110909
TEST TYPE: U
ALCO TARGET VALUE: .075
ALCO S/N: X172978

--- BREATH ANALYSIS ---

.075 ADJUSTED FOR 30.33 in		
ALCO TARGET	.076	10:11
BLANK TEST	.000	10:12
INTERNAL STANDARD	VERIFIED	10:12
ALCO TV 30.33 in	.076	10:12
BLANK TEST	.000	10:13
SUBJECT SAMPLE <i>COB 9/15/11</i>	.000	10:14
BLANK TEST	.000	10:14
ALCO TV 30.33 in	.076	10:15
BLANK TEST	.000	10:15

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130174

SEPTEMBER 09, 2011
TIME 10:16

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/08/09
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	30.35 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!@#%&'()*+,-./0123456789:;<=>?@BCDEFG
HIJKLMNPOQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrs tuvwxyz{|}~

COB 9/15/11