

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DEC 21 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130176</u>
Name		<u>DARYL REE</u>	ID# <u>2088</u> Date <u>12/09/09</u>
A	Agency	<u>SETKA POLICE DEPARTMENT</u>	Phone # <u>747-3245</u>
Instrument Location		<u>304 LAKE STREET</u>	
Alco S/N		<u>X124617</u>	Target Value <u>.085</u> High Pressure <u>400</u>
B	Alco Test Value Average	<u>.091</u>	<u>.091</u>
		1 st Alco	2 nd Alco
Signature		<u>Daryl Ree</u>	

(OVER)

AMB
12/21/09

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 7th day of May, 2009

Carolyn M. Noland (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130176

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130176
DECEMBER 09, 2009
OPERATOR'S NAME: RICE DARYL DEAN
OPERATOR'S NUMBER: 2088
SUBJECT'S LAST NAME: VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
UOC
O.L. #: UOC
DEPT/AGENCY: SITI
CASE/REPORT: UOC
TEST TYPE: U
ALCO TARGET VALUE: .085 ✓
ALCO S/N: X124617

--- BREATH ANALYSIS ---

.085 ADJUSTED FOR 30.55 in
ALCO TARGET .086 13:56
BLANK TEST .000 13:57
INTERNAL STANDARD VERIFIED 13:57
ALCO TV 30.55 in .091 13:57 ✓
BLANK TEST .000 13:58
SUBJECT SAMPLE .000 13:59
BLANK TEST .000 13:59
ALCO TV 30.55 in .091 14:00 ✓
BLANK TEST .000 14:00

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130176
DECEMBER 09, 2009
TIME 13:53
--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 01/08/09
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 30.55 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST
!*"#&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefg hijklmno
pqrs tuvwxyz{|}~