

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

APR 05 2010

Supervisor/Operator Performing the Verification Procedure: **DataMaster cdm S/N** 130176 ✓

Name DARYL RICE ID# 2088 Date 04/03/10 ✓

A Agency SETKA POLICE DEPT. Phone # 747-3245

Instrument Location 304 LAKE STREET ✓

B Alco S/N X301572 ✓ Target Value .078 ✓ High Pressure 1000

Alco Test Values .079 ✓ .080 ✓
1st Alco 2nd Alco

Signature Daryl Rice ✓

(OVER)

BMP
4/8/10


I, Nita J. Bolz, after being first duly sworn, depose and state as follows (to be read in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 26th day of April, 2010.

Carolyn M. Noland (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130176 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130176 ✓

APRIL 03, 2010 ✓

OPERATOR'S NAME:

RICE DARYL DEAN

OPERATOR'S NUMBER: 2998

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

UOC

O.L. #: UOC

DEPT/AGENCY: SIT1

CASE/REPORT: UOC

TEST TYPE: U

ALCO TARGET VALUE: .078 ✓

ALCO S/N: X301572 ✓

--- BREATH ANALYSIS ---

.078 ADJUSTED FOR 29.41 in ✓
ALCO TARGET .076 05:28
BLANK TEST .000 05:29
INTERNAL STANDARD VERIFIED 05:29
ALCO TV 29.42 .079 ✓ 05:29
BLANK TEST .000 05:30
SUBJECT SAMPLE .000 05:31
BLANK TEST .000 05:31
ALCO TV 29.42 in .000 05:32
BLANK TEST .000 05:32

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130176 ✓

APRIL 03, 2010 ✓

TIME 05:33

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 01/08/09

HEATERS

SAMPLE CHAMBER: 45c

BREATH TUBE: 42c

BAROMETER: 29.42 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKL#OPQRSTUWXYZ[\]^_`abcdefgijklmnop
qrstuvwxyzi{|}~