

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument  
State of Alaska

JUN 03 2010

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program


Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130176</u> ✓
A	Name <u>Bill Bandy</u>	ID# <u>7010</u> Date <u>5-31-10</u> ✓
	Agency <u>SITKA POLICE DEPT.</u>	Phone # <u>747-3245</u>
	Instrument Location <u>SITKA POLICE DEPT.</u> ✓	
B	Alco S/N <u>X301572</u> ✓	Target Value <u>.078</u> ✓ High Pressure <u>950</u>
	Alco Test Values <u>.082</u> ✓	<u>.082</u> ✓
	1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco
	Signature <u>Bill Bandy</u> ✓	<u>CB</u> <u>6/7/10</u>
	(OVER)	

I, Nita J. Bolz, after being first duly sworn, depose and state as follows: (Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 24<sup>th</sup> day of June, 2010.  
Carolyn M. Noland (Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130176 ✓

(CONTINUED FROM FRONT PAGE)

**C** TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130176 ✓

MAY 31, 2010 ✓

OPERATOR'S NAME:  
BANDY BILL

OPERATOR'S NUMBER: 7010

SUBJECT'S LAST NAME:

BANDY

SUBJECT'S FIRST NAME/MI :  
BILL

D.L. #: 6986272

DEPT/AGENCY: SIT1

CASE/REPORT:

TEST TYPE: U

ALCO TARGET VALUE: .078 ✓

ALCO S/N: X301572 ✓

--- BREATH ANALYSIS ---

.078 ADJUSTED FOR 30.07 in ✓  
ALCO TARGET .078 00:43  
BLANK TEST .000 00:44  
INTERNAL STANDARD (VERIFIED) 00:44  
ALCO TU 30.07 in .002 ✓ 00:45  
BLANK TEST .000 00:45  
SUBJECT SAMPLE *COB 6/7/10* .000 00:46  
BLANK TEST .000 00:47  
ALCO TU 30.07 in .002 ✓ 00:47  
BLANK TEST .000 00:48

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130176 ✓

MAY 31, 2010 ✓  
TIME 08:40

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 01/08/09  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 41c  
BAROMETER: 30.07 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

'###'()\*+,-./0123456789:;<=>@ABCDEFGHIJKL  
HIJKL#NOPQRSTUVWXYZ[\]^\_`abcdefgijklmnop  
qrstuvwxyz{|}~