

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130177 ✓
Name	Steve Kevan	ID#	6585 ✓
		Date	03/11/2011 ✓
A Agency	Alaska State Troopers	Phone #	(907) 438-2029
Instrument Location	St. Mary's Trooper Post		
B Alco S/N	X172959 ✓	Target Value	.081 ✓
		High Pressure	700
Alco Test Values	.081 ✓	.082 ✓	
	1 st Alco	2 nd Alco	
Signature			


(OVER)

AMB
3/11/11

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.




Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 1ST day of APRIL, 2011.




(Notary Seal Stamp)
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130177

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130177

MARCH 11, 2011

OPERATOR'S NAME:
KEHAN, STEVE
OPERATOR'S NUMBER: 6585
SUBJECT'S LAST NAME:
TEST
SUBJECT'S FIRST NAME/NI :
TEST
O.L. #: 00
DEPT/AGENCY: SMD
CASE/REPORT: 00
TEST TYPE: U
ALCO TARGET VALUE: .081
ALCO S/N: X172959

--- BREATH ANALYSIS ---

.081 ADJUSTED FOR 29.91 in
ALCO TARGET .080 11:18
BLANK TEST .000 11:19
INTERNAL STANDARD VERIFIED 11:19
ALCO TV 29.91 in .081 11:20
BLANK TEST .000 11:20
SUBJECT SAMPLE .080 11:21
BLANK TEST .000 11:22
ALCO TV 29.91 in .082 11:22
BLANK TEST .000 11:23

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130177

MARCH 11, 2011
TIME 11:09

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 4c
BREATH TUBE: 4c
BAROMETER: 29.91 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#&*()'+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefg|hijklmnop
qrstuvwxyz{|}~