

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

AUG 26 2009

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130177 ✓

Name STEVEN C. KEVAN ID# 6585 Date 8/25/09 ✓

A Agency ALASKA STATE TROOPERS Phone # 800-240-2019

Instrument Location ST. MARY'S ✓

B Alco S/N X301550 ✓ Target Value .079 ✓ High Pressure 700

Alco Test Value Average .081 ✓ .082 ✓
1st Alco 2nd Alco

Signature [Signature] ✓

CMB
8/24/09

(OVER)

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 29th day of Sept, 2009.

[Signature] (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130177 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130177

AUGUST 25, 2009 ✓ *COS 8/25/09*

OPERATOR'S NAME: *SCX 8/25/09*

KEVAN, STEVE

OPERATOR'S NUMBER: 6585

SUBJECT'S LAST NAME:

TEST

SUBJECT'S FIRST NAME/MI :

TEST

O.L. #: 00

DEPT/AGENCY: SMV8

CASE/REPORT: 00

TEST TYPE: U

ALCO TARGET VALUE: .079 ✓

ALCO S/N: X301550

BREATH ANALYSIS

.079 ADJUSTED FOR 29.83 in	✓	
ALCO TARGET	.078	10:13
BLANK TEST	.000	10:14
INTERNAL STANDARD	VERIFIED	10:14
ALCO TV 29.83 in	.081	10:15
BLANK TEST	.000	10:16
SUBJECT SAMPLE	.000	10:16
BLANK TEST	.000	10:17
ALCO TV 29.83 in	.082	10:17
BLANK TEST	.000	10:18

COS 8/25/09

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130177

AUGUST 25, 2009 ✓
TIME 10:19

DIAGNOSTIC CHECK

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/28/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 43c
BAROMETER: 29.85 in
FLOW DETECTOR: OKAY

PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789::;<=?@ABCDEF6
HIJKLMNPOQRSTUVWXYZ[\]^_`abcdefg|hijklmno
pqrstuvwxyz{|}~