



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130177
Name		Sean Adkins	ID# 5110 Date 10/06/2008 ✓
A	Agency	Alaska State Troopers	Phone # (907) 438-2029
Instrument Location		St. Mary's	✓
Alco S/N		X172966 ✓	Target Value .073 ✓ High Pressure 160 PSI ✓
B	Alco Test Value Average	.080 ✓ 1 st Alco	.078 ✓ 2 nd Alco
Signature			 10/24/08

(OVER)

(Do Not write in the area below)


Under the Alaska Rules of Evidence, I certify that;


(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Sean Adkins, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,





Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

10/24/08

Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130177 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130177

OCTOBER 06, 2008

OPERATOR'S NAME:

ADKINS/S/E

OPERATOR'S NUMBER: 5110

SUBJECT'S LAST NAME:

VERIFICATION OF CALIBRATION:

SUBJECT'S FIRST NAME/NI :

VOC

O.L. #: VOC

DEPT/AGENCY: SMD

CASE/REPORT: VOC

TEST TYPE: V

ALCO TARGET VALUE: .073 ✓

ALCO S/N: X172966 ✓

BREATH ANALYSIS

.073 ADJUSTED FOR 29.73 in
ALCO TARGET .072 09:58
BLANK TEST .000 09:59
INTERNAL STANDARD VERIFIED 09:59
ALCO TV 29.73 in .000 ✓ 10:00
BLANK TEST .000 10:00
SUBJECT SAMPLE .000 ✓ 10:01
BLANK TEST .000 ✓ 10:01
ALCO TV 29.73 in .078 ✓ 10:02
BLANK TEST .000 10:03

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130177

OCTOBER 06, 2008 ✓
TIME 09:55

DIAGNOSTIC CHECK

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 43c
BAROMETER: 29.73 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST
! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; (=) ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~