


# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130 178</u> ✓
Name	<u>James Kimura</u>	ID# <u>5677</u> Date <u>7/3/2008</u>
Agency	<u>Alaska State Troopers - Ketchikan Post</u> Phone # <u>(907) 225-5118</u>	
Instrument Location	<u>Ketchikan AST Post</u> ✓	
Alco S/N	<u>78820</u> ✓	Target Value <u>.077</u> ✓ High Pressure <u>950 PSI</u>
Alco Test Values	<u>.078</u> ✓ 1 <sup>st</sup> Alco	<u>.079</u> ✓ 2 <sup>nd</sup> Alco
Signature		 7/25/08

(OVER)


(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that;


(a) I am Jeane Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, James Kimura, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,



Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

7/28/08

Date

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130 178

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130178

JULY 03, 2008

OPERATOR'S NAME:

KIMURA/JAMES/K

OPERATOR'S NUMBER: 5677

SUBJECT'S LAST NAME:

UOC

SUBJECT'S FIRST NAME/MI :

UOC

O.L. #: 000000

DEPT/AGENCY: KETA

CASE/REPORT: 00-00000

TEST TYPE: U

ALCO TARGET VALUE: .077

ALCO S/N: 78820

--- BREATH ANALYSIS ---

.077 ADJUSTED FOR 29.94 in  
ALCO TARGET .077 01:45  
BLANK TEST .000 01:46  
INTERNAL STANDARD VERIFIED 01:46  
ALCO TU 29.94 in .078 01:47  
BLANK TEST .000 01:48  
SUBJECT SAMPLE .000 01:48  
BLANK TEST .000 01:49  
ALCO TU 29.94 in .079 01:49  
BLANK TEST .000 01:50

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130178

JULY 03, 2008

TIME 01:38

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 42c  
BAROMETER: 29.94 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z [ \ ] ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z { | } ~ #