

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

OCT 22 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130 178</u>	
Name <u>Trooper James K. Kimura</u>		ID# <u>5677</u>	Date <u>10/13/09</u>
Agency <u>Alaska State Troopers</u>		Phone # <u>(907) 225-5118</u>	
Instrument Location <u>Ketchikan AST Post</u>			
Alco S/N <u>X 124655</u>		Target Value <u>.078</u>	High Pressure <u>600 PSI</u>
Alco Test Values		<u>.078</u> 1 st Alco	<u>.081</u> 2 nd Alco
Signature <u>[Signature]</u>		CMB 10/13/09	

(OVER)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 6th day of NOV, 2009.

[Signature]
(Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130178 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130178 ✓

OCTOBER 13, 2009

OPERATOR'S NAME:

KIMURA/JAMES/K

OPERATOR'S NUMBER: 5677

SUBJECT'S LAST NAME:

UOC

SUBJECT'S FIRST NAME/NI :

UOC

O.L. #: 000000

DEPT/AGENCY: KETO

CASE/REPORT: 09-000000

TEST TYPE: U

ALCO TARGET VALUE: .078 ✓

ALCO S/N: X124655

--- BREATH ANALYSIS ---

.078 ADJUSTED FOR 29.46 in ✓
ALCO TARGET .076 18:01
BLANK TEST .000 18:02
INTERNAL STANDARD VERIFIED 18:02
ALCO TV 29.47 in .078 ✓ 18:03
BLANK TEST .000 18:03
SUBJECT SAMPLE *COS 10/13/09* .000 18:04
BLANK TEST .000 18:05
ALCO TV 29.47 in .081 ✓ 18:05
BLANK TEST .000 18:06

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130178 ✓

OCTOBER 13, 2009
TIME 18:06

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 29.47 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>@ABCDEFGHIJK
LMNOPQRSTUVWXYZ[\]^_`abcdefgijklmnop
qrstuvwxyz{|}~■