

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130179</u> ✓
Name		<u>John T. Ostoj</u>	ID# <u>4823</u> Date <u>10 MAY 08</u>
A	Agency	<u>Soldotna AST</u>	Phone # <u>262-4453</u>
Instrument Location		<u>Soldotna AST Post</u> ✓	
Alco S/N		<u>X 301596</u> ✓	Target Value <u>0.081</u> ^{0.081 CBS} ✓
			High Pressure <u>1000</u>
B	Alco Test Value Average	<u>.083</u> ✓	<u>.082</u> ✓
		1 st Alco	2 nd Alco
Signature		<u>[Signature]</u>	<u>CBS</u> <u>6/3/08</u>
(OVER)			

Under the Alaska Rules of Evidence, I certify that (Do not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, John Ostoj, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



[Signature]

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

6/5/08

Date

MAY 13 2008

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130179

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130179

MAY 10, 2008

OPERATOR'S NAME: *CS 6/3/08*
ESTRICH, JAVI
OPERATOR'S NUMBER: 4023
SUBJECT'S LAST NAME:
NOC

SUBJECT'S FIRST NAME: *CS 6/3/08*
TEST
CALIBRATION TEST
TEST METHOD: SLOW
USE/REPORT: TEST
SET TYPE: U
ALCO TARGET VALUE: .001
ALCO UNIT: 100/15%

--- BREATH ANALYSIS ---

.001 ADJUSTED FOR 29.81 in
ALCO TARGET .000 13:19
ALCO TEST .000 13:20
INTERNAL STANDARD VERIFIED 13:20
ALCO IN 29.81 in .000 13:20
BLOW TEST .000 13:21
SUBJECT SAMPLE *CS 6/3/08* .000 13:22
BLOW TEST .000 13:23
ALCO IN 29.81 in .000 13:23
BLOW TEST .000 13:24

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130179

MAY 10, 2008
TIME 13:30

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 29.81 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~