

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUL 29 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130179</u>
Name	<u>John T. Ostoj</u>	ID#	<u>4823</u> Date <u>7/27/09</u>
A Agency	<u>Alaska State Troopers</u>	Phone #	<u>262-4453</u>
Instrument Location	<u>Seldutna AST</u>		
B Alco S/N	<u>X124671</u>	Target Value	<u>.081</u> High Pressure <u>1200</u>
Alco Test Value Average	<u>.082</u>	<u>.083</u>	
	1 st Alco	2 nd Alco	
Signature	<u>[Signature]</u>		<u>COB</u> <u>7/30/09</u>
(OVER)			

I, Nita J. Bolz, after being first duly sworn, depose and state as follows. (Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 14th day of August, 2009.

[Signature] (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



