

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

010Z 4 0 91W
AUG 04 2010

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130179</u> ✓
Name	<u>James G. Truesdell</u>	ID# <u>2856</u> ✓ Date <u>8/1/10</u> ✓
A Agency	<u>Alaska State</u>	Phone # <u>262-4453</u>
Instrument Location <u>Soldotna AST</u>		
B Alco S/N	<u>78259</u> ✓	Target Value <u>0.083</u> ^{AMB} <u>0.084</u> ^{8/1/10} High Pressure <u>500</u>
Alco Test Value Average <u>.085</u> ✓ 1 st Alco 2 nd Alco <u>.084</u> ✓		
Signature <u>James G. Truesdell</u>		<u>COB</u> <u>8/9/10</u>
(OVER)		

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 11th day of AUGUST, 2010.

Daniel L. Nicholas (Notary Seal Stamp)
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130179

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130179

AUGUST 01, 2010

OPERATOR'S NAME: TRUESDELL/JAMES/G
OPERATOR'S NUMBER: 2956
SUBJECT'S LAST NAME: VERIFICATION
SUBJECT'S FIRST NAME/NI: OF/CAL

O.L. #: 1234567
DEPT/AGENCY: SLDM
CASE/REPORT:
TEST TYPE: U
ALCO TARGET VALUE: .083
ALCO S/N: 76259

--- BREATH ANALYSIS ---

.083 ADJUSTED FOR 30.38 in
ALCO TARGET .084 21:36
BLANK TEST .000 21:37
INTERNAL STANDARD VERIFIED 21:37
ALCO TU 30.38 in .085 21:37
BLANK TEST .000 21:38
SUBJECT SAMPLE ^{cab} 8/9/10 .000 21:39
BLANK TEST .000 21:39
ALCO TU 30.38 in .084 21:40
BLANK TEST .000 21:41

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130179

AUGUST 01, 2010
TIME 21:32

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 30.38 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

! "# \$ % ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~ ¡ ¢