

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130179 ✓

Supervisor/Operator Performing the Verification Procedure:

Name V. Eugene Fowler ID# 4386 ✓ Date 12-27-10 ✓

A Agency Alaska State Troopers Phone # 262-4453

Instrument Location Soldotna AST Post

B Alco S/N X301553 ✓ Target Value .081 ✓ High Pressure 200

Alco Test Value Average .079 ✓ .079 ✓
1st Alco 2nd Alco

Signature V. Fowler COB
1/18/11

(OVER)

Mailed
12/29/10
VFB


I, Nita J. Bolz, after being first duly sworn, depose and state as follows: (Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 2nd day of MAR, 2011 DNL
2010

Daniel L. Nicholas (Notary Seal Stamp)
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N L30179

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130179

DECEMBER 27, 2010

OPERATOR'S NAME:

FOWLER/A/E

OPERATOR'S NUMBER: 4386

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

D.L. #: VOC

DEPT/AGENCY: SLT5

CASE/REPORT: VOC

TEST TYPE: V

ALCO TARGET VALUE: .081

ALCO S/N: X301553

— BREATH ANALYSIS —

.081 ADJUSTED FOR 29.57 in
ALCO TARGET .080 19:03
BLANK TEST .080 19:04
INTERNAL STANDARD VERIFIED 19:04
ALCO TV 29.57 in .079 19:05
BLANK TEST .080 19:05
SUBJECT SAMPLE .080 19:06
BLANK TEST .080 19:07
ALCO TV 29.57 in .079 19:07
BLANK TEST .080 19:08

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130179

DECEMBER 27, 2010

TIME 19:00

— DIAGNOSTIC CHECK —

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/28/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 40c

BAROMETER: 29.57 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~ ¡ ¢