FEB 1 4 ~~2010~~ 2011 DW 3/17/11

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 30179

Supervisor/Operator Performing the Verification Procedure:

A

Name James Truesdell ID# 2856 Date 2/10/11

Agency Alaska State Troopers Phone # 262-4453

Instrument Location Soldotna ASI

B

Alco S/N 718493 Target Value .080 High Pressure 1100

Alco Test Value Average ~~.079~~ 0.077 BMB 3/6/11 .077

1st Alco 2nd Alco

Signature James J Truesdell

(OVER)

BMB 3/6/11

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
(3) I am the Scientific Director of the State Breath Alcohol Program.
(4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
(5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

STATE OF ALASKA SCIENTIFIC CRIME DETECTION LABORATORY

Subscribed and sworn before me this 17th day of MAR, ~~2010~~ 2011 DW

(Notary Seal Stamp)

Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office

DANIEL L. NICHOLAS Commission # 116780 NOTARY PUBLIC My Commission Expires with Office STATE OF ALASKA

BT9 03/00

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130179

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

ST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130179

FEBRUARY 10, 2011 BMB 3/6/11

OPERATOR'S NAME:
TRUESDELL/JAMES/G
OPERATOR'S NUMBER: 2856
SUBJECT'S LAST NAME:
VERIFICATION
SUBJECT'S FIRST NAME/MI :
OF/CAL
O.L. #: 1234567
DEPT/AGENCY: SL00
CASE/REPORT: 123
TEST TYPE: V
ALCO TARGET VALUE: .000
ALCO S/N: 718493

— BREATH ANALYSIS —

.000 ADJUSTED FOR 29.90 in

ALCO TARGET	.079	09:52
BLANK TEST	.000	09:53
INTERNAL STANDARD	VERIFIED	09:53
ALCO TV 29.90 in	.077	09:53
BLANK TEST	.000	09:54
SUBJECT SAMPLE	.000	09:55
BLANK TEST	.000	09:56
ALCO TV 29.90 in	.077	09:56
BLANK TEST	.000	09:57

BMB 3/6/11

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130179

FEBRUARY 10, 2011
TIME 09:47

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	41c
BAROMETER:	29.90 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~