

NOV 13 2011

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130180</u>
Name	<u>AW Anderson</u>	ID#	<u>10/26/11</u> Date <u>234-7640</u>
Agency	<u>Seldovia PD</u>	Phone #	_____
Instrument Location	<u>Seldovia PD</u>		
Alco S/N		Target Value	High Pressure
<u>X301530</u>		<u>.078</u>	<u>500 PSI</u>
Alco Test Values		1st Alco	2nd Alco
		<u>.073</u>	<u>.074</u>
Signature	<u>AW Anderson</u>		<u>CS</u> <u>11/4/11</u>

(OVER)

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 9 day of Nov, 2011.

Nikki Roth (Notary Seal Stamp)
 Daniel L. Nicholas
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

(CONTINUED FROM FRONT PAGE)

130180

COB 11/4/11

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130180

OCTOBER 26, 2011

COB
11/4/11

OPERATOR'S NAME:

ANDERSON AW

OPERATOR'S NUMBER: 2467

SUBJECT'S LAST NAME:

SIXTY DAY CALIBRATION

SUBJECT'S FIRST NAME/MI :

N/A

O.L. #: 1234567890

DEPT/AGENCY: 5001

CASE/REPORT: N/A

TEST TYPE: U

ALCO TARGET VALUE: .078

ALCO S/N: X301530

--- BREATH ANALYSIS ---

.078 ADJUSTED FOR 29.59 in		
ALCO TARGET	.077	16:03
BLANK TEST	.000	16:03
INTERNAL STANDARD	VERIFIED	16:03
ALCO TV 29.60 in	.073	16:04
BLANK TEST	.000	16:05
SUBJECT SAMPLE	.000	16:05
BLANK TEST	.000	16:06
ALCO TV 29.60 in	.074	16:07
BLANK TEST	.000	16:07

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130180

OCTOBER 26, 2011

TIME 16:00

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	43c
BREATH TUBE:	42c
BAROMETER:	29.60 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!##%&'()*+,-./0123456789:;<=>@ABCDEFGHIJK
LMNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxy z{|}~

