

*of Datamaster cdm
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Datamaster cdm S/N 130180

(CONTINUED FROM FRONT PAGE)

C **TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.**

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130180

JULY 30, 2009

*CAS
8/10/09*

OPERATOR'S NAME:
ANDERSON AW
OPERATOR'S NUMBER: 2467
SUBJECT'S LAST NAME:
SIXY DAY VERIFICATION
SUBJECT'S FIRST NAME/MI :
N/A
O.L. #: 1234567890
DEPT/AGENCY: SDU1
CASE/REPORT: N/A
TEST TYPE: U
ALCO TARGET VALUE: .078
ALCO S/N: X301532

--- BREATH ANALYSIS ---

.078 ADJUSTED FOR 30.01 in		
ALCO TARGET	.078	12:53
BLANK TEST	.000	12:54
INTERNAL STANDARD	VERIFIED	12:54
ALCO TU 30.01 in	.075	12:55
BLANK TEST	.000	12:55
SUBJECT SAMPLE	.000	12:56
BLANK TEST	.000	12:57
ALCO TU 30.01 in	.075	12:57
BLANK TEST	.000	12:58

*CAS
8/10/09*

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130180

JULY 30, 2009
TIME 12:50

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	30.01 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

```
!@#%&'()*+,-./0123456789:;<=>?@BCDEFG
HIJKLMNPOQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~
```

VERIFICATION OF CALIBRATION REPORT 1

of Datamaster cdm
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

AUG 03 2009

Supervisor/Operator Performing the Verification Procedure:	
Name <u>AW Anderson</u>	ID# <u>2467</u> Date <u>July 30, '09</u>
Agency <u>Seldovia Police Dept.</u>	Phone # <u>(907) 234-7640</u>
Instrument Location <u>Seldovia Police Dept</u>	
Alco S/N <u>X301530</u> Target Value <u>.078</u> High Pressure <u>500 PSI</u>	
Alco Test Value Average <u>.075</u>	<u>.075</u>
1 st Alco	2 nd Alco
Signature <u>AW Anderson</u>	<u>CRS</u> <u>8/6/09</u>

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 14th day of August, 2009.

Carolyn M. Noland (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office

