VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N 130180

Supervisor/Operator Performing the Verification Procedure:

Name AW Anderson ID# 2467 Date 11/24/08

A

Agency Seldovia Police Dept. Phone # (907) 234-7640

Instrument Location Seldovia Police Dept.

B

Alco S/N X 301530 Target Value .078 High Pressure 100 PSI

Alco Test Value Average .073 (1st Alco) .073 (2nd Alco) 12/2/08

Signature AW Anderson

(OVER)

(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, A W Anderson, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

12/2/08

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

Date

BT9 01/00

NOV 28 2008

VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N 130180

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130180

NOVEMBER 24, 2008

OPERATOR'S NAME:
ANDERSON, AW
OPERATOR'S NUMBER: 2467
SUBJECT'S LAST NAME:
VERIFICATION,60 DAY
SUBJECT'S FIRST NAME/MI :
N/A
O.L. #: 1234567890
DEPT/AGENCY: SOV1
CASE/REPORT: N/A
TEST TYPE: V
ALCO TARGET VALUE: .078
ALCO S/N: X301530

--- BREATH ANALYSIS ---

.078 ADJUSTED FOR 29.02 in

ALCO TARGET	.075	09:50
BLANK TEST	.000	09:51
INTERNAL STANDARD	VERIFIED	09:51
ALCO TV 29.02 in	.073	09:52
BLANK TEST	.000	09:52
SUBJECT SAMPLE	.000	09:53
BLANK TEST	.000	09:54
ALCO TV 29.02 in	.073	09:54
BLANK TEST	.000	09:55

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130180

NOVEMBER 24, 2008
TIME 09:46

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	29.02 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~