

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

FEB 09 2012

DataMaster cdm S/N <u>130181</u>		
Supervisor/Operator Performing the Verification Procedure:		
Name <u>Randy Simmen</u>	ID# <u>5123</u> Date <u>1-26-12</u>	
A Agency <u>AST</u>	Phone # <u>328-0130</u>	
Instrument Location <u>Nenah</u>		
B Alco S/N <u>78895</u> Target Value <u>.074</u> High Pressure <u>500</u>		
Alco Test Values	<u>.078</u> 1st Alco	<u>.077</u> 2nd Alco
Signature <u>[Signature]</u>	<u>AST</u>	<u>COS</u> <u>2/9/12</u> <u>COS</u> <u>2/9/12</u>

(OVER)

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 5 day of April, 2012.

[Signature] (Notary Seal Stamp)
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130181

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130181

JANUARY 26, 2012 *COB 2/9/12*

OPERATOR'S NAME: SIMMONSONNYA
OPERATOR'S NUMBER: 5123
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
A
O.L. #: 1
DEPT/AGENCY: NENO
CASE/REPORT: 1
TEST TYPE: V
ALCO TARGET VALUE: .074
ALCO S/N: 78895

--- BREATH ANALYSIS ---

.074 ADJUSTED FOR 29.94 in		
ALCO TARGET	.074	23:51
BLANK TEST	.000	23:52
INTERNAL STANDARD	VERIFIED	23:52
ALCO TV 29.94 in	.078	23:53
BLANK TEST	.000	23:53
SUBJECT SAMPLE <i>COB 2/9/12</i>	.060	23:54
BLANK TEST	.000	23:54
ALCO TV 29.94 in	.077	23:55
BLANK TEST	.000	23:56

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130181

JANUARY 26, 2012
TIME 23:57

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01

HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c

BAROMETER: 29.94 in

FLOW DETECTOR: OKAY

PUMP
HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST
!"#\$%&'()*+,-./0123456789:;<=?@ABCDEFGHI
HIJKLMNOPQRSTUVWXYZ\]^_`abcdefgijklmnop
pqrsuvwxyz{|}~