

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130181</u> ✓
Name <u>Malik Jones</u>		ID# <u>5830</u> Date <u>1/29/09</u> ✓
A	Agency <u>AST</u>	Phone # <u>451-5706</u>
Instrument Location <u>Nenana</u> ✓		
Alco S/N <u>X301532</u> ✓ Target Value <u>.079</u> ✓ High Pressure <u>400</u>		
Alco Test Value Average <u>.078</u> ✓ <u>.078</u> ✓		
1 st Alco 2 nd Alco		
Signature  ✓		 ✓
(OVER)		


Under the Alaska Rules of Evidence, I certify that;
(Do Not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Malik Jones, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

2/9/09
Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130181

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130181

JANUARY 29, 2009

OPERATOR'S NAME: JONES/MALIK/JAMAL
OPERATOR'S NUMBER: 5830
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI:
V

D.L. #: 1
DEPT/AGENCY: NENI
CASE/REPORT:
TEST TYPE: V
ALCO TARGET VALUE: .079 ✓
ALCO S/N: X301532

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 29.46 in		
ALCO TARGET	.077	12:07
BLANK TEST	.000	12:08
INTERNAL STANDARD	VERIFIED	12:08
ALCO TV 29.46 in	.078 ✓	12:08
BLANK TEST	.000	12:09
SUBJECT SAMPLE	.000	12:10
BLANK TEST	.000	12:11
ALCO TV 29.47 in	.078 ✓	12:11
BLANK TEST	.000	12:12

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130181

JANUARY 29, 2009
TIME 12:13

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 40c
BAROMETER: 29.47 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST
!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKLMNPOQRSTUVWXYZ[\]^_`abcde fghi jk lmnop
qrs tuvwxyz{|}~"