# VERIFICATION OF CALIBRATION REPORT

*of DataMaster cdm Breath Test Instrument*
*State of Alaska*
*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

**DataMaster cdm S/N** 130181

Supervisor/Operator Performing the Verification Procedure:

**A**

Name JSwartz ID# 4314 Date 3/15/08

Agency SCDL Phone # 269-5592

Instrument Location SCDL

**B**

Alco S/N Scott07 Target Value .080 High Pressure 600 psi

Alco Test Values .078 (1st Alco) .077 (2nd Alco)

Signature [signature] CDS 3/19/08

(OVER)

(Do not write in the area below)

Under the Alaska Rules of Evidence, I certify that:

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Jeanne M Swartz, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

[signature] 3/20/08

Jeanne Swartz Date
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

BT9 06/03

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster cdm breath Test Instrument*
*State of Alaska*
*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

**DataMaster cdm S/N** 130181

(CONTINUED FROM FRONT PAGE)

**C** TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

**Breath Test Record**

Serial Number 130181 Model DataMaster
Location ANC1 Anchorage Police Dept.

CDB 3/19/08

**Subject Information**

Last Name Verification Of Cal O.L. # L
First Name L

**Operator Information**

Operator Name SwartzJ Case
Operator # 4314 Department AAA1

**Test Results**

| | | | | | |
|---|---|---|---|---|---|
| Alco Target | .080 | Alco Serial | SCOTT07 | | |
| Nominal | 0.079 | Barometer | 29.91 | Test Type | V |
| Dry Gas 1 | 0.078 | Time: | 11:11:29 | Test Date | 03/15/2008 |
| Subject Alcohol | 0.000 | Time: | 11:13:11 | Test Time | 11:10:09 |
| Dry Gas 2 | 0.077 | Time: | 11:14:17 | Report | Y |

CDB 3/19/08

DIAGNOSTIC CHECK

**Diagnostic Test Record**

Serial Number 130181 Model DataMaster
Location ANC1 Anchorage Police Dept.

**Diagnostic Test Results**

| | | | |
|---|---|---|---|
| Test Date | 03/15/2008 | Computer | Okay |
| Test Time | 11:06:53 | Program | Okay |
| | | Software Date | 02/20/2001 |
| | | Sample Chamber (C) | 48 |
| | | Breath Tube (C) | 43 |
| | | Barometer (in) | 29.91 |
| | | Flow Detector | Okay |
| | | High Speed Pump | Okay |
| | | Detector | Okay |
| | | Filter | Okay |
| | | Quartz | Okay |
| | | Calibration | Okay |