

**VERIFICATION OF CALIBRATION REPORT**  
*of Datamaster cdm*  
*State of Alaska*  
**Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program**

Supervisor/Operator Performing the Verification Procedure: <u>Datamaster cdm S/N 130181 ✓</u>	
Name <u>Michael A Wery</u>	ID# <u>3590</u> Date <u>9-27-08 ✓</u>
A Agency <u>AST</u> Phone # <u>451-5100</u>	
Instrument Location <u>Nenana Post</u>	
B Alco S/N <del>X130181</del> <sup>X301532</sup> <sub>11/14/08</sub> <sup>COB</sup> Target Value <u>.079 ✓</u> High Pressure <u>900</u>	
Alco Test Value Average <u>.078 ✓</u>	<u>.077 ✓</u>
1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco
Signature <u>Michael A. Wery ✓</u>	
(OVER)	COB 11/24/08

Under the Alaska Rules of Evidence, I certify that:  
(Do Not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Michael A Wery, whose name appears in my records, is qualified to verify the calibration of the above instrument

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody



Signed and sealed,

Jeanne Swartz

Jeanne Swartz  
 Scientific Director Designee  
 State of Alaska  
 State Breath Alcohol Testing Program

11/21/08  
 Date

OCT 03 2008

# VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N 130181 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130181

SEPTEMBER 27, 2008

OPERATOR'S NAME:

WERY MICHAEL/A

OPERATOR'S NUMBER: 3590

SUBJECT'S LAST NAME:

VERIFICATION

SUBJECT'S FIRST NAME/MI:  
OF CALIBRATION

O.L. #: 0

DEPT/AGENCY: MEMO

CASE/REPORT:

TEST TYPE: V

ALCO TARGET VALUE: .079 ✓

ALCO S/N: ~~130181~~ X801532 ✓

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 29.67 in

ALCO TARGET	.078 ✓	13:41
BLANK TEST	.000	13:42
INTERNAL STANDARD	VERIFIED	13:42
ALCO TV 29.67 in	.079 ✓	13:42
BLANK TEST	.000	13:43
SUBJECT SAMPLE	.000	13:43
BLANK TEST	.000	13:44
ALCO TV 29.67 in	.077 ✓	13:44
BLANK TEST	.000	13:45

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130181

SEPTEMBER 27, 2008 ✓

TIME 13:35

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 42c

BAROMETER: 29.67 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#%&'()\*+,-./0123456789:;<=>?@ABCDEF  
HIJKLNPQRSTUWXYZ[\]^\_`abcde/ghijklmno  
pqrsuvwxyz{|}~