

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

MAY 12 2009

Supervisor/Operator Performing the Verification Procedure:	
DataMaster cdm S/N <u>130183</u> ✓	
Name	<u>HUTT, William T.</u> ID# <u>3359</u> Date <u>05/12/09</u> ✓
A Agency	<u>Homer Police Dept. / Jail</u> Phone # <u>235-3150</u>
Instrument Location	<u>4060 Heath St. Homer, AK 99603</u> ✓
Alco S/N <u>75260</u> ✓ Target Value <u>.079</u> ✓ High Pressure <u><50 psi</u>	
Alco Test Value Average <u>.086</u> ✓ <u>.086</u> ✓	
1 st Alco 2 nd Alco	
Signature	<u>William T. Hutt</u> ✓
<u>083</u> <u>5/18/09</u>	

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 27th day of May, 2009.

Carolyn M. Noland (Notary Seal Stamp)

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130183 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130183

MAY 12, 2009 ✓ *CSB*

OPERATOR'S NAME:
MUTT, WILLIAM T

OPERATOR'S NUMBER: 3359

SUBJECT'S LAST NAME:
VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :
ETOH

O.L. #: 01234567

DEPT/AGENCY: HOMI

CASE/REPORT: 09-0000

TEST TYPE: U

ALCO TARGET VALUE: .079 ✓

ALCO S/N: 75260 ✓

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 29.91 in		
ALCO TARGET	.078	23:14
BLANK TEST	.000	23:15
INTERNAL STANDARD	VERIFIED	23:15
ALCO TV 29.91 in	.086 ✓	23:16
BLANK TEST	.000	23:16
SUBJECT SAMPLE	.000	23:17
BLANK TEST	.000	23:18
ALCO TV 29.91 in	.086 ✓	23:19
BLANK TEST	.000	23:19

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130183

MAY 12, 2009 ✓ *WTHA*
TIME 23:10

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 41c

BAROMETER: 29.91 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST
!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJKL
MNOPQRSTUVWXYZ[\]^_`abdefghijklmnop
qrstuvwxyz{|}~@