

barometric pressure
30.03

VERIFICATION OF CALIBRATION REPORT JAN 03 2010
of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130183

Supervisor/Operator Performing the Verification Procedure:

Name HUTT, William T. ID# 3359 Date 12/31/09

A Agency Homer Police Dept./Jail Phone # 235-3150

Instrument Location 4060 Heath St. Homer, AK 99603 235-3151 (fax)

B Alco S/N X301599 Target Value .086 High Pressure 350 psi

Alco Test Values .088 .087
1st Alco 2nd Alco

Signature William T. Hutt

(OVER)

BMB
1/13/10


I, Nita J. Bolz, after being first duly sworn, depose and state as follows. (Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 26th day of January, 2010.

Carolyn M. Noland (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 132183

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130183

DECEMBER 31, 2009

OPERATOR'S NAME: *BMB 11/31/0*
MUTT, WILLIAM T
OPERATOR'S NUMBER: 3359
SUBJECT'S LAST NAME:
IDENTIFICATION OF CAL
SUBJECT'S FIRST NAME: *WTHR*
SEX:
AGE:
WEIGHT:
HEIGHT:
HAIR:
EYES:
EYES: 03-0000
TEST TYPE: 0
ALCOHOL TARGET VALUE: .086
ALCOHOL S/N: X301599

--- BREATH ANALYSIS ---

.086 ADJUSTED FOR 29.96 in
ALCOHOL TARGET .086 01:26
BLANK TEST .000 01:27
INTERNAL STANDARD VERIFIED 01:27
ALCOHOL TO 29.96 in .086 01:27
BLANK TEST .000 01:29
SUBJECT SAMPLE *WTHR* .086 01:30
BLANK TEST .000 01:31
ALCOHOL TO 29.96 in .087 01:31
BLANK TEST .000 01:32

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130183

DECEMBER 31, 2009
TIME: 01:21

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 00/00/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 29.96 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

11/11/00 11:11:11 11/11/00 11:11:11 11/11/00 11:11:11
11/11/00 11:11:11 11/11/00 11:11:11 11/11/00 11:11:11
11/11/00 11:11:11 11/11/00 11:11:11 11/11/00 11:11:11