

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130184

Supervisor/Operator Performing the Verification Procedure:
Name JOHN HAMILTON ID# 3192 Date 4/10/12

A Agency PETERSBURG POLICE DEPARTMENT Phone # 772-3838
Instrument Location PETERSBURG

B Alco S/N 4L Target Value .080 High Pressure 500

Alco Test Values .078 .078
1st Alco 2nd Alco

Signature [Signature]

(OVER)

BMB
4/20/12

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 14 day of June, 2012 NAK 6/14/12

[Signature] (Notary Seal Stamp)
Daniel L. Nicholas - Nikki Roth 6/14/12
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130184

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130184

APRIL 10, 2012

OPERATOR'S NAME:
HAMILTON JOHN A
OPERATOR'S NUMBER: 3192
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :

A
O.L. #: A
DEPT/AGENCY: PSG1
CASE/REPORT: A
TEST TYPE: V
ALCO TARGET VALUE: .000
ALCO S/N: 4L

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 29.90 in
ALCO TARGET .079 13:43
BLANK TEST .000 13:44
INTERNAL STANDARD VERIFIED 13:44
ALCO TU 29.91 in .078 13:44
BLANK TEST .000 13:45
SUBJECT SAMPLE .000 13:46
BLANK TEST .000 13:47
ALCO TU 29.90 in .078 13:47
BLANK TEST .000 13:48

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130184

APRIL 10, 2012
TIME 13:40

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 29.91 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmnop
qrstuvwxyz{|}~■