

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:	
Name <u>TODD BRUND</u>	ID# <u>7672</u> Date <u>3/24/12</u>
Agency <u>GLACIER BAY NATIONAL PARK</u>	Phone # <u>697 2687</u>
Instrument Location <u>BARNETT COVE, AK</u>	
Alco S/N <u>2K</u>	Target Value <u>0.081</u> High Pressure <u>650 PSI</u>
(ADJUSTED)	
Alco Test Values	<u>0.081</u> <u>0.081</u>
1st Alco	2 nd Alco
Signature <u>Todd Brund</u>	

(OVER)

BMB
3/30/12

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 5 day of April, 2012.

Nikki Roth (Notary Seal Stamp)
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130185

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130185

MARCH 24, 2012

OPERATOR'S NAME:

BRUND/TODD/M

OPERATOR'S NUMBER: 7572

SUBJECT'S LAST NAME:

TEST

SUBJECT'S FIRST NAME/MI :

TEST

O.L. #: 0000

DEPT/AGENCY: GUS3

CASE/REPORT: 0000

TEST TYPE: T

ALCO TARGET VALUE: 0.352

ALCO S/N: 2K

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 30.37 in
ALCO TARGET .001 18:25
BLANK TEST .000 18:26
INTERNAL STANDARD VERIFIED 18:26
ALCO TV 30.37 in .001 18:26
BLANK TEST .000 18:27
SUBJECT SAMPLE .352 18:27
BLANK TEST .000 18:28
ALCO TV 30.37 in .001 18:28
BLANK TEST .000 18:29

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130185

MARCH 24, 2012

TIME 18:28

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 01/08/09

HEATERS

SAMPLE CHAMBER: 42c

BREATH TUBE: 42c

BAROMETER: 30.37 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: 0:57

CALIBRATION: OKAY

PRINTER TEST

HT:ALCO TV 30.37 in
pqrstuvwxyz(1)*****

