

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130185

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130185

DECEMBER 03, 2007 *COB 12/13/07*

OPERATOR'S NAME:
CHRISTENSEN, S, E
OPERATOR'S NUMBER: 1315
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
VOC
O.L. #: 12/03/07
DEPT/AGENCY: JNU1
CASE/REPORT: 12/03/07
TEST TYPE: V
ALCO TARGET VALUE: .075
ALCO S/N: X172968

--- BREATH ANALYSIS ---

.075 ADJUSTED FOR 29.98 in
ALCO TARGET .075 05:09
BLANK TEST .000 05:10
INTERNAL STANDARD VERIFIED 05:10
ALCO TV 29.96 in .078 05:11
BLANK TEST .000 *COB 12/13/07* 05:11
SUBJECT SAMPLE .000 05:13
BLANK TEST .000 05:13
ALCO TV 29.96 in .079 05:14
BLANK TEST .000 05:15

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130185

DECEMBER 03, 2007
TIME 05:15

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 48c
BREATH TUBE: 42c
BAROMETER: 29.96 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF
HIJKLMN OPQRSTUVWXYZ[\]^_`abcdefghijklmnop
qrstuvwxyz{|}~