



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130185

(CONTINUED FROM FRONT PAGE)

**C** TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130185

JANUARY 12, 2008

OPERATOR'S NAME:  
CHRISTENSEN, S, E  
OPERATOR'S NUMBER: 1315  
SUBJECT'S LAST NAME:  
VERIFICATION OF CAL  
SUBJECT'S FIRST NAME/MI :  
VOC

O.L. #: 011208  
DEPT/AGENCY: JN01  
CASE/REPORT: 011208  
TEST TYPE: V  
ALCO TARGET VALUE: .075  
ALCO S/N: X172968

--- BREATH ANALYSIS ---

.075 ADJUSTED FOR 29.75 in		
ALCO TARGET	.074	00:30
BLANK TEST	.000	00:31
INTERNAL STANDARD	VERIFIED	00:31
ALCO TV 29.75 in	.078	00:31
BLANK TEST	.000	00:32
SUBJECT SAMPLE	.000	00:33
BLANK TEST	.000	00:33
ALCO TV 29.75 in	.078	00:34
BLANK TEST	.000	00:34

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130185

JANUARY 12, 2008  
TIME 00:35

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS:	
SAMPLE CHAMBER:	49c
BREATH TUBE:	41c
BAROMETER:	29.75 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST  
!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHIJKL  
HIJKLMNOPQRSTUVWXYZ[\]^\_`abcdefgijklmno  
pqrstuvwxyz{|}~\*~