VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130185

Supervisor/Operator Performing the Verification Procedure:

Name BRANDON LAWRENSON ID# 66009 Date 07/15/08

A. Agency JUNEAU POLICE DEPT Phone # 586-0600

Instrument Location JUNEAU PD

Alco S/N X75602 Target Value .077 High Pressure 900

B. COB 8/18/08

Alco Test Values .079 (1st Alco) .079 (2nd Alco)

Signature COB 8/18/08

(OVER)

(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz , Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Brandon Lawrenson , whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed, 8/20/08

Jeanne Swartz Date
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

AUG 14 2008

BTD 03/00

JUL 18 2008

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130185

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130185

JULY 15, 2008

OPERATOR'S NAME:
LAWRENSON/BRANDON/J
OPERATOR'S NUMBER: 6609
SUBJECT'S LAST NAME:
VOC
SUBJECT'S FIRST NAME/MI :
VOC
O.L. #: 0000000
DEPT/AGENCY: JNU1
CASE/REPORT: 012345
TEST TYPE: V
ALCO TARGET VALUE: .077
ALCO S/N: 475602

--- BREATH ANALYSIS ---

.077 ADJUSTED FOR 30.30 in

ALCO TARGET	.077	14:00
BLANK TEST	.000	14:01
INTERNAL STANDARD	VERIFIED	14:01
ALCO TV 30.30 in	.079	14:01
BLANK TEST	.000	14:02
SUBJECT SAMPLE	.000	14:02
BLANK TEST	.000	14:03
ALCO TV 30.30 in	.079	14:03
BLANK TEST	.000	14:04

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130185

JULY 15, 2008
TIME 13:57

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	30.30 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~